



## CE DECLARATION OF CONFORMITY

<b>Instrument Model:</b>	APEX (all models)	
<b>Manufacturer:</b>	Cardinal Scale Manufacturing Company Detecto Scale 203 East Daugherty Street Webb City, Missouri 64870 USA Telephone No. + 417 673 4631 Fax No. + 417 673 5001	
<b>Type Examination Certificate(s):</b>	755 SPZ300	DK0199.558 R60/2000-NL1-17.42 Rev1
<b>Test Certificate(s):</b>	755 SPZ300	DANAK 1916180 R60/2000-NL1-10.20/Nmi- 10200947-06
<b>Notified Body Issuing Approval for Module B: EU Type Examination (Annex II.2 of 2014/31/EU):</b>	0199-Delta (755) 0122-NMi (SPZ300)	

Conformity to the following Directives is demonstrated by reference to the following harmonized standards or technical documentation

<u>Applicable Directives</u>	<u>Harmonized Standards or Other Technical Documentation</u>
<b>Directive on Non-Automatic Weighing Instruments (2014/31) including all Essential Requirements of Annex I:</b>	EN45501:2015
<b>Directive on Electromagnetic Compatibility (2014/35):</b>	EN61000-4-2 for Electrostatic discharge immunity and EN61000-4-3 for Radiated, radio frequency, and electromagnetic field immunity test.
<b>Directive on Electrical Equipment designed to use within certain Voltage limits (2006/95):</b>	EN61000-4-7 for General guide on harmonics measurements and instrumentation for power supply systems and connected equipment.
<b>Medical Device Directive (93/42/EEC)</b>	ISO13485:2016

The Technical Construction File required by this Directive is maintained at the corporate headquarters of Cardinal Scale Manufacturing Company.

This declaration is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of Cardinal Scale Manufacturing 203 East Daugherty, Webb City MO, 64870:

Mark Levels  
Manager, Quality Assurance

27<sup>th</sup> of September 2018  
Rev C