

# Adscope® Sprague

## Instructions for Use



Thank you for choosing an ADC® Adscope® Stethoscope. Your new stethoscope is designed to provide unsurpassed acoustic sensitivity, unrivalled durability, and unparalleled comfort. Every component has been carefully designed to maximize performance.

### Device Description and Intended Use

This device is intended for medical diagnostic purposes only. It can be used for listening to heart, lung and other body sounds as well as for normal auscultation. The device is not designed, sold, or intended for use except as indicated.

### Symbol Definitions

The following symbols are associated with your stethoscope.

Symbol	Definition	Symbol	Definition
	Important Caution		Date of Manufacture
	Not made with natural rubber latex		Batch Code
	Phthalate Free		Catalog Number
	Consult instructions for use		Medical Device
	Meets the general safety and performance requirements of Regulation (EU) 2017/745 of the European Union		Unique Device Identifier
	Authorized representative in the European Community/European Union		Non-Sterile
	Manufacturer		Do not use if the package is damaged
			Importer
			Distributor

### General Warnings

**A warning statement in this manual identifies a condition or practice which, if not corrected or discontinued immediately, could lead to patient injury, illness, or death.**

- WARNING:** Never use stethoscope without eartips firmly locked in place.
  - WARNING:** Stethoscope contains small parts which may present a choking hazard.
  - WARNING:** Cross-contamination of the stethoscope could cause patient infection. Clean and disinfect stethoscope between uses as described in Care and Maintenance.
  - WARNING:** Eartips can be swallowed and cause a choking hazard. Ensure all parts are properly attached and stored.
  - WARNING:** Stethoscope tubing can be a strangulation hazard. Keep away from children.
- NOTE:** When first unpacked, your Sprague tubing may experience some kinks from packaging. This is normal. To uncurl, hang scope for 24 hours at room temperature.

### Chestpiece Options

The ADC Sprague design utilizes five different interchangeable chestpiece attachments (three bells and two diaphragms) to detect a broad frequency range in both pediatric and adult patients. The chestpiece drum accommodates two different chestpiece attachments of your choice at one time. Select the type that best suits the intended patient application.



**Adult diaphragm assembly** - for filtering out lower frequencies and accenting the higher ranges in adults.


**Pediatric diaphragm assembly** - for detecting high pitched murmurs in infants and children.

**1 1/4" bell** - for detecting low and medium pitched sounds and murmurs in adults.

**1" bell** - for detecting low and medium pitched sounds and murmurs in children.

**3/4" bell** - for detecting low and medium pitched sounds and murmurs in infants.


Chestpiece attachments can be affixed by threading clockwise until tight. To remove, twist in a counterclockwise direction. (Figure 1)



### Selection of the Chestpiece

On ADC Adscope Spragues, the entire chestpiece rotates 180°, allowing for the selection of one side at a time. Only one side is operable at a time. ADC's patented valve mechanism eliminates virtually all acoustic leakage from the inoperable side.

To select the side in operation, hold the chestpiece stem in one hand and rotate the chestpiece with the other hand until a click is felt and heard. A light tap on the diaphragm or bell will help determine which side is in operation as well.



Adult Diaphragm Activated



Pediatric Diaphragm Activated

### Chestpiece Key

ADC's proprietary chestpiece design and slotted key allows you to tighten the pediatric or adult assembly as securely as you like (to reduce risk of loss) without fear of being unable to disassemble the rim in order to replace a worn or broken diaphragm (a problem that often plagues less costly Sprague scopes).




### Replacing the Diaphragm

To replace a diaphragm, remove the desired assembly by gripping the assembly firmly and turning counterclockwise.



To loosen the diaphragm rim, insert the key into the two slots on the threaded portion, located on the underside of the assembly. While holding the rim steady, turn the key counterclockwise to loosen the rim.

Remove worn diaphragm and replace it with a new one. Reassemble the rim and tighten by turning key clockwise until tight. Then reassemble diaphragm rim back onto chestpiece drum by turning clockwise until it is tight.




### Orientation of the Headset

To obtain the most comfortable, acoustically sealed aural fit, rotate each binaural tube approximately 15° forward, towards the bridge of your nose. Simply twist each tube where it enters the tubing.



### Adjusting the Spring Tension in the Binaurals


To reduce the tension, grip the binaurals firmly using both hands with fingers at the base of the binaural, thumbs at each side just before the binaurals meet the tubing. Flex binaural spring gradually until desired tension is obtained. To increase the tension, grip the binaurals in the hand and squeeze the binaural tubes together gradually.



**Caution:** Excessive adjustment of the spring tension could weaken the spring.

### Replacing Adsoft™ Snap-on Eartips


Slide Adsoft Snap-on eartip over adapter unit it locks tight into internal channel. Repeat steps for second aural tube.



### Care and Maintenance

- Avoid extreme heat, cold, solvents, or oils.
- Do not leave exposed to direct sunlight for prolonged periods of time.
- Binaurals and tubing can be wiped down with 70% isopropyl alcohol solution or mild soapy water.
- Eartips may be removed from binaurals for a thorough cleaning.
- Periodically check that eartips are securely affixed to the headset and free from damage.
- Do not immerse stethoscope in any liquid.
- Do not steam sterilize. If sterilization is necessary, gas sterilize only.

### Replacement Parts List



Ref. No.	Item#	Part Name	Qty.	Ref. No.	Item#	Part Name	Qty.
1	640-01D	Chestpiece Drum Assembly	1	11,12	602-078BK1Q	Adsoft Snap-On Eartips, Large & Small	2 pr.
2	640-02S	Retaining Ring, Ped.	1	13	640-11	Bell, 1 1/4"	1
3	640-03S	Diaphragm, Ped.	1	14	640-12	Bell, 1"	1
4	640-15	Zinc Disk, Ped.	1	15	640-13	Bell, 3/4"	1
5	640-14	Zinc Disk, Adult	1	16	640-18	Accessory Pouch	1
6	640-03	Diaphragm, Adult	1	17	640-19	Chestpiece Key	1
7	640-02	Retaining Ring, Adult	1	3	640-03S	Diaphragm, Ped. (spare)	1
8	641-05T	Tubing, 22" (spec. color)	2	6	640-03	Diaphragm, Adult (spare)	1
9	640-16	Tubing Clip	2	-	640-06	Tubing Caps	2
10	640-05	Binaurals with Lg Eartips	1				

### LIMITED WARRANTY

Please register your stethoscope online at: <http://www.adctoday.com/register>. ADC warrants its products against defects in materials and workmanship under normal use and service as follows:

1. Warranty service extends to the original retail purchaser only and commences with the date of delivery.
2. Your Adscope Sprague stethoscope is warranted for five years from date of purchase (all parts).

**What Is Covered:** Replacement of parts, and labor.

**What Is Not Covered:** Transportation charges to and from ADC. Damages caused by abuse, misuse, accident, or negligence. Incidental, special, or consequential damages. Some states do not allow the exclusion or limitation of incidental, special, or consequential damages, so this limitation may not apply to you.

**To Obtain Warranty Service:** Send item(s) postage paid to ADC, Attn: Repair Dept., 55 Commerce Dr., Hauppauge, NY 11788. Please include your name and address, phone no., proof of purchase, and a brief note explaining the problem.

**Implied Warranty:** Any implied warranty shall be limited in duration to the terms of this warranty and in no case beyond the original selling price (except where prohibited by law). This warranty gives you specific legal rights and you may have other rights which vary from state to state.

Any serious incident that has occurred in relation to this medical device should be reported to ADC and the competent authority of the Member State in which the user and/or patient is established.



Inspected in the U.S.A.  
Made in China

