

INSTRUCTIONS FOR USE

# D920<sup>MKIII</sup> & D930<sup>MKIII</sup>

Handheld Dopplers



# **INSTRUCTIONS FOR USE D920/D930**

## **Contents**

- 1. Introduction**
  - 1.1 Unpacking / Preliminary Checks**
- 2. Safety**
  - 2.1 Warnings**
  - 2.2 Patient Applied Parts**
  - 2.3 Intended Use & Indications**
  - 2.4 Contraindications**
  - 2.5 Patient Population**
- 3. Warranty & Service**
  - 3.1 Service Life**
  - 3.2 Maintenance and Repair**
- 4. Product Identification**
  - 4.1 Product Controls**
  - 4.2 Symbol Identification**
- 5. Prepare the Doppler for Use**
  - 5.1 Battery Insertion / Replacement**
- 6. Operation**
  - 6.1 Battery Saver**
  - 6.2 After Use**
- 7. Care and Cleaning**
  - 7.1 General Care**
  - 7.2 Cleaning and Disinfecting the Doppler**
  - 7.3 Cleaning and Disinfecting Probes**
- 8. Specifications**
  - 8.1 Equipment Classification**
  - 8.2 General**
  - 8.3 Environmental**
  - 8.4 Standards Compliance**
- 9. Electromagnetic Compatibility**

# 1. Introduction

The D920 & D930 are battery powered, hand-held Dopplers for obstetric use. They provide an audible fetal heart sound. This equipment is for use only by suitably qualified healthcare practitioners and is not intended for use by the patient.

Before using this equipment, study this manual carefully and familiarise yourself with the controls, display features and operation.

Experience with use of ultrasonic dopplers is preferable, but for novice users training material is provided with the online documents. Exposure to ultrasound should be kept As Low As Reasonably Achievable - (ALARA guidelines). This is considered to be good practice and should be observed at all times.

Scan the QR code on the rear cover of this IFU with a smartphone, or visit the Huntleigh website for electronic copies of user literature. All documents are available to download as PDF files. To read them, you must have a PDF reader installed on your device. Alternatively paper copies are available upon request.

## 1.1 Unpacking / Preliminary Checks

On receipt of your Doppler, check that all items are present and undamaged. If items are missing or have been damaged in transit, inform Huntleigh Healthcare immediately.

|                  |                     |           |
|------------------|---------------------|-----------|
| Handheld Doppler | IFU (this document) | Batteries |
| Ultrasound Gel   |                     |           |

# 2. Safety

## 2.1 Warnings

- Dopplers are screening tools to aid the healthcare professional. If there is doubt as to fetal status, further investigations should be undertaken immediately using alternative techniques.
- Always ensure that Maternal HR or any Artefacts are not misinterpreted as fetal heartrate. If in doubt, feel the mother's pulse during the examination.
- Do not use in the presence of flammable gases.
- Do not use in a sterile field unless additional barrier precautions are taken.
- Do not sterilise the product or its accessories. The product will be damaged.
- Do not expose to excessive heat, including prolonged exposure to sunlight.
- Do not dispose of batteries in fire as this can cause them to explode.
- The Doppler is not waterproof and must not be immersed.
- This product contains sensitive electronics, which are susceptible to interference, this will be indicated by unusual sounds.
- This equipment must not be modified.

## 2.2 Patient Applied Parts

As defined in IEC 60601-1, the patient applied parts of the Dopplers are the ultrasound probes.

## 2.3 Intended Use & Indications

The Dopplers are intended for use by qualified healthcare practitioners in primary, acute and community healthcare environments, for the assessment of fetal heart rate..

They are indicated for routine screening of pregnant women of all ages from early gestation through to full term, and for low risk labour management.

## 2.4 Contraindications

- Do not use on broken or fragile skin.
- Do not use on the eye.

## 2.5 Patient Population

The D920 and D930 are suitable for use on all patient populations.

## 3. Warranty & Service

Huntleigh Healthcare Diagnostic Products Division standard terms and conditions apply to all sales. A copy is available on request. These contain full details of warranty terms and do not limit the statutory rights of the consumer.

**Service Returns:** To return the Doppler:

- Clean the product following the instructions in this manual.
- Pack it in suitable packing.
- Attach a decontamination certificate (or other statement declaring that the product has been cleaned) to the outside of the package. (Huntleigh Healthcare Ltd reserve the right to return product that does not contain a decontamination certificate).
- Mark the package 'Service Department'.

### 3.1 Service Life

This is defined as the period during which the device is expected to remain safe and suitable to meet its intended use, and all risk control measures remain effective.

The service life for this device is seven years.

### 3.2 Maintenance and Repair

There are no user serviceable parts inside the Doppler unit or probes. This product does not require periodic maintenance. Inspection is recommended each time the product is used, paying particular attention to the tip of the probes, checking for cracks etc., and to the cable. Any unusual sounds or intermittent behaviour should be investigated.

Spare parts are available. Please refer to service manual for further information and part numbers. A full technical description is provided in the Service Manual 793329.


## Caution

---


**Servicing cannot be carried out while the Doppler is in use.**

## 4. Product Identification

### 4.1 Product Controls



|    |                                            |
|----|--------------------------------------------|
| 1  | Headphone Socket                           |
| 2  | Green LED: indicates power ON              |
| 3  | Yellow LED: flashes when batteries are low |
| 4  | Loudspeaker                                |
| 5  | Probe Holder                               |
| 6  | Trolley Mount                              |
| 7  | On/Off/Volume                              |
| 8  | Pocket Clip                                |
| 9  | Battery Compartment                        |
| 10 | Rear Panel Label                           |



## 4.2 Symbol Identification


|                                                                                     |                                                                                                                                                                                                                                              |                                                                                     |                                                                  |
|-------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------|------------------------------------------------------------------|
|    | Patient applied parts (ultrasound probes) are type BF according to the definitions in IEC 60601-1.                                                                                                                                           |                                                                                     |                                                                  |
|    | General Warning                                                                                                                                                                                                                              |    | Attention, consult accompanying documents / Instructions for Use |
|    | This symbol signifies that this product, including its accessories and consumables is subject to the WEEE (Waste Electrical and Electronic Equipment) regulations and should be disposed of responsibly in accordance with local procedures. |                                                                                     |                                                                  |
|    | This symbol signifies that this product complies with the essential requirements of the Medical Device Directive (93/42/EEC) - Medical Device Regulation (EU/2017/745).                                                                      |                                                                                     |                                                                  |
| RxOnly                                                                              | Federal law restricts this device to sale by, or on the order of a licensed healthcare practitioner.                                                                                                                                         |                                                                                     |                                                                  |
|                                                                                     |                                                                                                                                                                                                                                              |                                                                                     |                                                                  |
|    | Legal Manufacturer in association with the CE mark in Europe<br>ArjoHuntleigh AB<br>Hans Michelsensgatan 10 211 20 Malmö, Sweden                                                                                                             |                                                                                     |                                                                  |
| IP20                                                                                | Protected against ingress of solid foreign objects >12.5 mm diameter.<br>Not protected against ingress of water.                                                                                                                             |                                                                                     |                                                                  |
| IPx7                                                                                | Protected against ingress of water 1 m immersion for up to 30 minutes.                                                                                                                                                                       |                                                                                     |                                                                  |
|  | Power On/Off                                                                                                                                                                                                                                 |  | Battery                                                          |
|  | Device Identifier                                                                                                                                                                                                                            |  | Serial Number                                                    |
|  | Reference Number                                                                                                                                                                                                                             |  | Medical Device                                                   |
|  | Fragile                                                                                                                                                                                                                                      |  | Keep Dry                                                         |
|  | Atmospheric Pressure Limitations                                                                                                                                                                                                             |  | Relative Humidity Limitations                                    |
|  | Temperature Limitations                                                                                                                                                                                                                      |  | Cardboard packaging can be recycled                              |
|  | LATEX FREE<br>Does not contain Latex                                                                                                                                                                                                         |  | PVC FREE<br>Does not contain PVC                                 |
|  | Headphone Socket                                                                                                                                                                                                                             |  | Volume                                                           |

**Note:** Product labelling should be readable from a distance of up to 0.7m.


## 5. Prepare the Doppler for Use

### 5.1 Battery Insertion / Replacement

Disconnect the Doppler from other equipment before removing the battery cover.



Insert a suitable tool into the recess, release the clip and gently lever off the battery cover.



Insert the batteries according to the diagram, observing polarity.

- Use only alkaline LR6 (AA) non-rechargeable batteries.

**Note:** If the Doppler will not be used for an extended period, remove the batteries.

#### Low Battery Indication

The yellow LED flashes when the batteries are low. The batteries should then be replaced for reliable operation.

## 6. Operation

Turn the On/Off/Volume control to the On position. The green LED indicates power On.

**Note:** D920 is fitted with a 2 MHz probe while D930 is fitted with a 3 MHz probe. The probes and cables are waterproof but do NOT immerse the Doppler hand unit.

#### Clinical Use

Make the patient comfortable in a semi-recumbant or sitting position. Apply a liberal amount of gel to the abdomen. Place the faceplate of the probe flat against the abdomen above the symphysis pubis. Adjust the probe to get an optimum audio signal, ideally by angling the probe around while maintaining firm pressure. Avoid sliding the probe over the skin. Best performance is from the fetal heart itself, characterised by 'slapping' valve sounds rather than umbilical artery or placental sounds.

### 6.1 Battery Saver

To prolong battery life, the Doppler will automatically go to sleep after five minutes of continuous operation.

To wake the Doppler, turn the On/Off/Volume control to the Off position, then On again.

### 6.2 After Use

Turn the On/Off/Volume control to the Off position.

Refer to the cleaning section before storing or using the unit on another patient.

## 7. Care and Cleaning

### 7.1 General Care

The Doppler contains delicate components, particularly the probe tip, which should be handled and treated with care. Periodically, and whenever the integrity of the system is in doubt, carry out a check of all functions as described in this IFU. If there are any defects, contact Huntleigh or your distributor for repair or to order a replacement.

#### Caution

- Check with your facility's local infection control policy and medical equipment cleaning procedures.
- Observe warnings and guidance on cleaning fluid labelling regarding use and personal protective equipment (PPE).
- If detergent or disinfectant wipes are used, make sure that excess solution is squeezed from the wipe prior to use.
- Always switch off the Doppler before cleaning and disinfecting.
- Always wipe off disinfectant using a cloth dampened with clean water.
- Do not allow any fluid to enter the products and do not immerse in any solution.
- Do not use abrasive cloths or cleaners.
- Do not use automatic washers or autoclaves.
- Do not use Phenolic detergent based disinfectants, solutions containing cationic surfactants, ammonia based compounds or perfumes and antiseptic solutions.

### 7.2 Cleaning and Disinfecting the Doppler

Always keep the external surfaces clean and free of dirt and fluids using a clean dry cloth.

- Wipe any fluids from the surface of the product using a clean dry cloth.
- Wipe with a cloth dampened in 70% Isopropyl Alcohol.
- Completely dry with a clean, dry cloth.
- If the product has been contaminated use the methods described for probes.

### 7.3 Cleaning and Disinfecting Probes

Clean the probes before examining a patient using the low risk cleaning method below. Following patient examination, clean and/or disinfect the probes by the appropriate method based upon the level of cross contamination risk, as defined below:

| Risk   | Definitions                                                                                   | Procedure                                                                                                                                                                                     |
|--------|-----------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Low    | Normal use or low risk situations include patients having intact skin and no known infection. | 1. Remove soiling, wipe with a mild neutral detergent and then wipe with a cloth dampened in water.<br>2. Completely dry with a clean cloth.                                                  |
| Medium | The patient has a known infection, skin is not intact, the part is heavily soiled.            | 1. Follow low risk procedure then wipe with a cloth dampened in Sodium Hypochlorite (1,000ppm).<br>2. After two minutes wipe with a cloth dampened in water and then dry with a clean cloth.  |
| High   | This procedure should only be used when the part has been contaminated by blood.              | 1. Follow low risk procedure then wipe with a cloth dampened in Sodium Hypochlorite (10,000ppm).<br>2. After two minutes wipe with a cloth dampened in water and then dry with a clean cloth. |

#### Caution

**Repeated and unnecessary use of concentrated solutions will result in damage to the product. Do not allow Sodium Hypochlorite solutions to come into contact with metal parts.**

The use of disinfectant materials other than those listed is the responsibility of the user for their efficacy and compatibility with the device.

## 8. Specifications

### 8.1 Equipment Classification

|                                                                                                                                  |                                                                                                                              |
|----------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------|
| Type of protection against electric shock                                                                                        | Internally powered equipment                                                                                                 |
| Degree of protection against electric shock<br> | Type BF - equipment with an applied part  |
| Mode of operation.                                                                                                               | Continuous                                                                                                                   |
| Degree of protection against harmful ingress of particles and/or water.                                                          | Main Unit: IP20*,<br>Probes : IPX7                                                                                           |
| Degree of safety of application in the presence of a flammable anaesthetic                                                       | Equipment not suitable for use in the presence of a FLAMMABLE ANAESTHETIC MIXTURE WITH AIR, OXYGEN OR NITROUS OXIDE          |

\*For home use, this can be upgraded to IPX2 when using the protective pouch (ACC-OBS-080).

### 8.2 General

|                       |                                                                            |        |       |
|-----------------------|----------------------------------------------------------------------------|--------|-------|
| Max. Audio Output     | 500mW rms typical (loudspeaker)                                            |        |       |
| Auto shut-off         | After five minutes of continuous operation                                 |        |       |
| Headphone output      | Max. output Power: 25 mW rms (32Ω)<br>Connector: 3.5 mm stereo jack socket |        |       |
| Battery Type and Life | LR6 (AA) Alkaline cells, 2 x 1.5 V, Typically 500 one minute examinations  |        |       |
| Size                  | 140 x 33 x 75 mm                                                           | Weight | 267 g |

### 8.3 Environmental

|                                         |                             |
|-----------------------------------------|-----------------------------|
| Operating                               |                             |
| Temperature range                       | +5°C to +40°C               |
| Relative Humidity                       | 15% to 90% (non condensing) |
| Pressure                                | 700 hPa to 1060 hPa         |
| Transport and Storage between uses      |                             |
| Without relative humidity control       | -25°C to +5°C               |
| At a r.h. of up to 90% non-condensing   | +5°C to +35°C               |
| At a water vapour pressure up to 50 hPa | >+35°C to +70°C             |

### 8.4 Standards Compliance

|                                                                                                            |                      |
|------------------------------------------------------------------------------------------------------------|----------------------|
| IEC 60601-1: 2020                                                                                          | IEC 60601-1-11: 2015 |
| IEC 60601-2-37: 2015 Thermal Indices (TI) and Mechanical Index (MI) are below 1.0 for all device settings. | IEC 60601-1-2: 2014  |

## 9. Electromagnetic Compatibility

Make sure the environment in which the Doppler is installed is not subject to strong sources of electromagnetic interference (e.g. radio transmitters, mobile phones). This equipment generates and uses radio frequency energy. If not installed and used properly, in strict accordance with the manufacturer's instructions, it may cause or be subject to interference. Type-tested in a fully configured system, complies with IEC 60601-1-2, the standard intended to provide reasonable protection against such interference. Whether the equipment causes interference may be determined by turning the equipment off and on. If it does cause or is affected by interference, one or more of the following measures may correct the interference:


- Reorient the equipment
- Relocate the equipment with respect to the source of interference
- Move the equipment away from the device with which it is interfering

### Warnings

- The use of accessories, transducers and cables other than those specified, with the exception of transducers and cables sold by the manufacturer of the Doppler as replacement parts for internal components, may result in increased emissions or decreased immunity of the Doppler.
- The Doppler should not be used adjacent to or stacked with other equipment and that if adjacent or stacked use is necessary, the Doppler should be observed to verify normal operation in the configuration in which it will be used.
- Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the Doppler including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.

This section is only applicable to United Kingdom (UK) market when UK marking is applied to the Arjo medical device labelling.

#### UK Symbol:



#### UK Responsible Person & UK Importer:

**Arjo (UK) Ltd., ArjoHuntleigh House, Houghton Regis. LU5 5XF**

Is the appointed UK Responsible Person as defined in UK Medical Devices Regulations 2002 (SI 2002 No 618, as amended).

For Northern Ireland (NI) CE marking will still apply until further amendment to applicable regulations.

EN

If a serious incident occurs in relation to this medical device, affecting the user, or the patient then the user or patient should report the serious incident to the medical device manufacturer or the distributor.  
In the European Union, the user should also report the serious incident to the Competent Authority in the member state where they are located.

Registered No: 942245 England & Wales.  
©Huntleigh Healthcare Limited 2021

**A Member of the Arjo Family**

As our policy is one of continuous improvement, we reserve the right to modify designs without prior notice.

# HUNTLEIGH

1001048-4