

## **Autoclave Sterilization for Titanium NonMagnetic Surgical Instruments**

<b>TEMPERATURE</b>	<b>PRESSURE</b>	<b>TIME</b>
132° c (270° F)	20 PSI	3 Minutes

Indicator tape, autoclave bags etc should be used as instructed by your facility guidelines.